

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130493</u>	
Name <u>PATRICK SMITH</u>	ID# <u>6397</u> Date <u>16 JUN 08</u>
A Agency <u>CLEAN SECURITY FORCES</u> ✓ Phone # <u>585-6313</u>	
Instrument Location <u>BLD 6 37 CLEAN AFS</u>	
B Alco S/N <u>X301548</u> ✓ Target Value <u>-078</u> High Pressure <u>1100</u>	
Alco Test Values <u>.080</u> <u>-078</u> 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u> <u>COB</u> <u>7/7/08</u>	
(OVER)	

Under the Alaska Rules of Evidence, I certify that: (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Patrick J Smith, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

[Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

7/8/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130493

JUNE 16, 2008

OPERATOR'S NAME:

SMITH/PATRICK/J

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 00000000

DEPT/AGENCY: CLR4

CASE/REPORT: 00000000

TEST TYPE: V

ALCO TARGET VALUE: .078

ALCO S/N: X301548

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.52 in

ALCO TARGET .076 20:26

BLANK TEST .000 20:27

INTERNAL STANDARD VERIFIED 20:27

ALCO TV 29.52 in .000 20:27

BLANK TEST .000 20:28

SUBJECT SAMPLE .000 20:29

BLANK TEST .000 20:29

ALCO TV 29.52 in .078 20:30

BLANK TEST .000 20:31

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130493

JUNE 16, 2008

TIME 20:23

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 29.54 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~